




CuringPen

LED Curing Light USER MANUAL

Changzhou Sifary Medical Technology Co.


Content

1. Scope of CuringPen	4
1.1 Parts Identification	4
1.2 Components and accessories	5
1.3 Optional accessories	5
2. Symbols Used	6
3. Before Use	7
3.1 Scope of application	7
3.2 Contraindications	7
4. Setting up the CuringPen	8
4.1 Install the light source head	8
4.2 Install the disposable sleeve	8
4.3 Install the protective light shield	8
4.4 Charging	10
5. Use Interface	11
5.1 Panel key	11
6. Setting	12
6.1 Selecting memory	12
6.2 Advanced setting	13
7. Operation	14
7.1 Charge	14
7.2 Handpiece operation	15
7.3 Operation mode	17
8. Maintenance	19
9. Error Warning	20
10. Troubleshooting	21
11. Technical Data	23
12. EMC Tables	24
13. Statement	29


1. Scope of CuringPen

1.1 Parts Identification

- ,1Light source head
- ,2 Protective light shield
- ,3Disposable sleeve
- ,4handpiece
- ,5Adapter
- ,6Charging base



1.2 Components and accessories

-handpiece (1pcs) ORDER CODE:6551003 	Light source head(1pcs) ORDER CODE:6551001 
Disposable sleeve (100pcs) ORDER CODE:6542002 	Adapter (1pcs) ORDER CODE:6516001 
	



Charging base (1pcs)

Protective light shield (1pcs)

Błąd! Użyj karty Narzędzia główne, aby zastosować
标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główn
e, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

ORDER CODE:6551004



ORDER CODE:6551005

1.3 Optional accessories

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główn e, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

Light curing depth test board (1pcs) ORDER CODE:6503011	Protective light glasses (1pcs) ORDER CODE:6541001
--	---

2. Symbols Used

 Warning	If the instructions are not followed properly, operation may lead to hazards for the product or the user/patient.
 NOTE	Additional information, explanation of operation and performance.
	Serial number
	Catalogue number
	Manufacturer
	Date of manufacture
	Safety class II device
	Type B applied part
	Direct current
	WEEE(Waste of Electrical and Electronic Equipment)Directive Marking
	Store in a dry place
	Consult instructions for use
	Authorized Representative in the European Community
	Temperature limitation
	Relative humidity
	Transport and storage pressure conditions:70 kPa - 106 kPa
	Manufacturer's LOGO

3. Before Use

3.1 Scope of application

Suitable for the purpose of curing dental resins and composites or detecting caries.

This device must only be used in hospital environments, clinics or dental offices by qualified dental personnel

3.2 Contraindications

Do not use the device for non-endodontic dental procedures.

Safety and effectiveness have not been established in pregnant women and children.



Warning

Read the following warnings before use:

1. The device must not be placed in humid surroundings or anywhere where it can come into contact with any type of liquids.
2. Do not expose the device to direct or indirect heat sources. The device must be operated and stored in a safe environment.
3. The device requires special precautions with regard to electromagnetic compatibility (EMC) and must be installed and operated in strict compliance with the EMC information. In particular, do not use the device in the vicinity of fluorescent lamps, radio transmitters, remote controls, portable or mobile RF communication devices and do not use this system near the active HF Surgical Equipment in the hospital .Do not charge, operate or store at high temperatures. Comply with the specified operating and storage conditions.
4. Protective light shield and a disposable sleeve are compulsory during treatment.
5. If irregularities occur in the device during treatment, switch it off. Contact the agency.
6. Never open or repair the device yourself, otherwise, void the warranty

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

4. Setting up the CuringPen

4.1 Install the light source head

Make sure the light source head alignment the slots of handpiece, plug them together until it "click" securely into place.



The light source head can be 360 degrees rotated without take off, make it easy to watch the LCD in treatment




WARNING

- After connecting the light source head and the handpiece, pull it gently to make sure the connection is good.
- Make sure the assembly is connected properly, otherwise might cause unexpected fault, even hurt the patients


4.2 Install the disposable sleeve

Take a disposable sleeve and gently insert it into the handpiece from the light source head.



4.3 Install the protective light shield

Make sure the light source head alignment the slots of the protective light shield, plug them together.



WARNING


- Disposable sleeve must be discarded after use
- The light source head, protective light shield and handpiece should be cleaned and disinfected after each treatment.

Avoid direct exposure of the lamp to the eyes during working

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

4.4 Charging


Plug the USB of adapter into the charge base, and plug the other end into a power outlet, the Power LED on charge base will light up (green).



NOTE

- Only the original adapter could be used.

Put the handpiece all the way into the charge base, the charge state will show on the screen.



Charging indication appears on the screen, and flashes slowly , when battery is fully charged or in a state near full charge, the flash will stop





NOTE

- Put the handpiece into the charge base in the right direction, otherwise the handpiece will not be recharged.

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

5.Use Interface



5.1 Panel key

	○ ,1🔘 main switch ○ ,2“S” Setting key Turn Power On Press 🔘 more than 0.5 seconds to turn on the instrument Memory Change Press setting key during standby state Turn Power Off Long press 🔘 to turn power off. Advanced setting During power off state, holding down press .“S” then press 🔘 to entry advanced setting, Press “S” till target setting, press 🔘 to adjust, then press “S”to confirm
	○ ,1 Battery ○ ,2 Memory ○ ,3 Curing light intensity ○ ,4 Curing time

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

6.Setting

6.1 Selecting memory

	Memory Change There are 6 built-in memory programs, namely M0, M1, M2, RAMP, PULSE, Detect. Press setting key to change the memory during standby state										
	Light intensity setting The memory programs of M0, M1 and M2 can be customized to 2300mW/cm², 1500mW/cm² or 1000mW/cm². Long Press “S” to enter the light intensity setting menu, then press “S” to change or press main switch to confirm and exit. On the light intensity setting menu, long press “S” to enter the time setting menu.  NOTE • The light intensity of RAMP, PULSE and Detect memory programs are built-in, and the user cannot modify the settings.										
	Curing Time setting In the curing time setting menu, press “S” to select different times. The time selection is different under different light intensity: <table><tr><th>light intensity(mW/cm²)</th><th>time selection(sec)</th></tr><tr><td>2300</td><td>01,02,03</td></tr><tr><td>1500</td><td>05,10,15,20</td></tr><tr><td>1000</td><td>05,10,15,20</td></tr><tr><td>600</td><td>10,15,20,25</td></tr></table> Long press “S” or short press to exit the setting. The current memory program automatically saves the setting parameters.	light intensity(mW/cm ²)	time selection(sec)	2300	01,02,03	1500	05,10,15,20	1000	05,10,15,20	600	10,15,20,25
light intensity(mW/cm ²)	time selection(sec)										
2300	01,02,03										
1500	05,10,15,20										
1000	05,10,15,20										
600	10,15,20,25										

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

6.2 Advanced setting

	During power off state, holding down press “S” then press  to entry the “Auto power off time” setting. It can be set to 1min, 5min, 10min or 15min. Press  to change setting, press “S” to enter the next setting “Volume” volume setting mode.
	In the volume setting menu, press “S” to select different settings. It can be set to “Off”, “low”, “mid” or “high”. Press  to change setting, press “S” to enter the next setting “Hand” hand setting mode.
	The right hand and the left hand can be set in the “Hand” setting menu, Press  to change setting, press “S” to enter the next setting.
	The last item is the firmware version, and the handpiece shuts down automatically within 1 s after the version information is displayed.


Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

7.Operation

7.1 Charge

	Displays the present remaining amount of the battery. Less than 15% remains, please charge.  NOTE ● If the power is less than 15%, must be recharged within 30 days, otherwise the battery will be damaged.
	Charging indication appears on the screen, and flashes slowly, when battery is fully charged or in a state near full charge, the flash will stop Fully charged will take about 4 hours, depending on residual battery power and battery state. It can be recharged 300-500 times, depending on the operating conditions of the device.  NOTE ● When charging, other function will forcibly stop, take from charge base, press main switch, the last function will recall.
 WARNING ● Do not change the battery, only trained technician or distributor can change the battery, the electronic parts will be damaged if use a wrong battery or install with a wrong way.	

7.2 Handpiece operation



After the handpiece main unit is turned on, press  to start the light intensity output, and the time will start counting down. Press any key to interrupt the output during the light intensity output.

There is a beep every 5s while working, and the output is automatically turned off after the countdown ends.



WARNING

- When the device is working, do not directly illuminate the eyes, otherwise it will cause injury.
- Do not directly illuminate the skin, otherwise high temperature burns may occur.



NOTE



- After the device is used continuously for 10 times at 2300mW/cm^2 , it will be prohibited to use the highest light intensity output within 1 minute, Used to prevent the lamp from overheating
- When using, the light should be directly irradiated onto the curing dental resins and

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.



	composites to avoid improper exposure. ● The disposable protective cover and protective light shield are highly recommended.
	When the device is in standby mode, it will automatically shut down when it reaches the set time. The factory default setting is 5 minutes
WARNING ● Before using, please try it outside the oral cavity to ensure that there is no problem with the function of the device. ● Do not disassemble the lamp during treatment .	
NOTE ● If there is any abnormal functioning, stop using the device and report to company ● Gloves are compulsory during treatment ● Always clean the handpiece and light source head after each treatment.	

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

7.3 Operation mode


M0、M1、M2	Standard mode In these three modes, the light intensity can be set to 2300mW/cm², 1500mW/cm² or 1000mW/cm². Wavelength: 385nm-515nm When  pressed, the setting light intensity is output immediately, and the sound is prompted every 5 seconds. 
RAMP	Ramp Mode Light intensity : 1000mW/cm². Wavelength: 385nm-515nm When  pressed, the light intensity gradually increases to 1000mW/cm² in the first 5 seconds, and then continues to output 1000mW/cm², prompting the sound every 5 seconds. 
PULSE	PULSE Mode Light intensity : 1000mW/cm²

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

	Wavelength: 385nm-515nm When ☉ pressed, the output light intensity is 1000mW/cm² and flashes once every 1 second, prompting the sound every 5 seconds.  The graph shows 'Light intensity' on the y-axis and 'Time' on the x-axis. A horizontal line at 100% intensity is shown with a series of vertical pulses, indicating periodic flashing.
Detect	Detecting mode Light intensity : 600mW/cm² Wavelength: 385nm-410nm  The graph shows 'Light intensity' on the y-axis and 'Time' on the x-axis. A horizontal line at 100% intensity is shown with a single vertical pulse, indicating a single flash. When ☉ pressed, the setting light intensity is output immediately, and the sound is prompted every 5 seconds.  NOTE - This mode is only suitable for detecting dental calculus, caries, cracks, etc.  WARNING - Detect Mode cannot be used to cure dental resins and composites, otherwise there is a risk of incomplete curing.

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.






8.Maintenance

Disinfection components		
handpiece 	Light source head 	Adapter 
Charging base 	Protective light shield 	
Wipe all the surfaces with a cloth lightly moistened with Ethanol for Disinfection (Ethanol 70 to 80vol%) at least 2min, repeat for 5 times.		
WARNING - Do not use anything except Ethanol for Disinfection (Ethanol 70 to 80 vol%). - Do not use too much ethanol as it's going into machine and damage the components inside.		

Disposable components
Disposable sleeve (100pcs) 
WARNING - Please discard it after use and do not use it again. - The product has been disinfected with ethylene oxide, so there is no need to disinfect it before use.

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

9.Error Warning

	The battery is too low. Charge it immediately
	Blue LED broken, contact your distributor.
	Blue LED overload, contact your distributor.
	Violet LED broken, contact your distributor.
	Violet LED overload, contact your distributor.
	The temperature of the light source head is higher than expectation, turn the power off or waiting more than 1 minutes to let it cold down.

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

10.Troubleshooting

When trouble is found, check the following points before contacting your distributor. If none of these are applicable or the trouble is not remedied even after action has been taken, the product may have failed. Contact your distributor.

Problem	Cause	Solution	Ref ch ap
The power is not turned on.	The battery is run down.	Charge the battery.	7.1
	The time to press the main switch is too short.	Press the main switch more than 0.5 seconds.	5.1
The Power LED does not light up when charging.	Use a wrong adapter.	Use the original adapter.	/
	There is no electricity in the outlet.	Check the connection.	/
	The adapter is not connected.	Check the connection.	/
the handpiece does not display the charging interface while charging	Put the handpiece into the charge base in the wrong direction.	Check the direction.	4.4
	The charge base broken.	Using adapter connect to handpiece directly, and contact your distributor.	/
	Charge pin of charge base unable to rebound.	Remove debris which between move part and base of the charge pin.	/
No output	Handpiece broken	Contact your distributor	/
No sound	Beep volume set to off	Set beep volume to low, mid or high.	/
	Handpiece broken	Contact your distributor	
Insufficient light intensity	There are resin or other contaminants on the surface of the lamp lens	Cleaning the lamp head residue	

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

11. Technical Data

Manufacturer	Changzhou Sifary Medical Technology Co.,Ltd
Model	CuringPen
Dimensions	22cm x 11cmx 8cm±1cm
Weight	730g ±10%
Power supply	Lithium ion battery: 3.7V, 1600mAh, ±10%
Charger power supply	AC 100-240 V, ±10%
Charger power output	5V  1A
Power Frequency	50/60Hz, ±10%
Charger nominal power input	5VA
Light intensity	2300mW/cm ² 1500mW/cm ² 1000mW/cm ² 600 mW/cm ²
Wavelength	385nm-515nm
Electrical safety class	Class II
Applied part	B
Operation conditions	Use: in enclosed spaces Ambient temperature: 5°C ~ 40 °C Relative humidity: <80% Operating altitude < 3000m above sea level Atmospheric pressure: 70kPa ~ 106kPa
Transport and storage conditions	Ambient temperature: -20 °C ~ +55 °C Relative humidity: 20% ~ 80 % Atmospheric pressure: 70kPa ~ 106kPa

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

12.EMC Tables

Guidance and manufacturer’s declaration – electromagnetic emissions		
The CuringPen is intended for use in the electromagnetic environment specified below. The customer or the user of the CuringPen should assure that it is used in such an environment.		
Emissions test	Compliance	Electromagnetic environment-guidance
RF emissions CISPR 11	Group 1	The CuringPen uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class B	The CuringPen is suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes.
Harmonic emissions IEC61000-3-2	Class A	
Voltage fluctuations/flicker emissions IEC 61000-3-3	Complies	

Guidance and manufacturer’s declaration – electromagnetic immunity			
The CuringPen is intended for use in the electromagnetic environment specified below. The customer or the user of the CuringPen should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment – guidance
Electrostatic discharge (ESD) IEC 61000-4-2	+/- 8 kV contact +/- 2 kV, +/- 4 kV, +/- 8 kV, +/- 15 kV air	+/- 8 kV contact +/- 2 kV, +/- 4 kV, +/- 8 kV, +/- 15 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd!
Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

			should be at least 30 %.
Electrical fast Transients /bursts IEC 61000-4-4	±2kV 100kHz repetition frequency	±2kV 100kHz repetition frequency	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	Line to line: ±0.5kV, ±1kV Line to earth: ±0.5kV, ±1kV, ±2kV	Line to line: ±0.5kV, ±1kV Line to earth: ±0.5kV, ±1kV, ±2kV	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply lines IEC 61000-4-11	0% UT; 0.5 cycle at 0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315° 0% UT; 1 cycle and 70% UT; 25/30 cycles sine phase at 0° 0% UT; 250/300 cycle	0% UT; 0.5 cycle at 0°, 45°, 90°, 135°, 180°, 225°, 270°, and 315° 0% UT; 1 cycle and 70% UT; 25/30 cycles sine phase at 0° 0% UT; 250/300 cycle	Mains power quality should be that of a typical commercial or hospital environment. If the user of devices require continued operation during power mains interruptions, it is recommended that devices be powered from an uninterruptible power supply or a battery

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd!
Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

Power frequency magnetic field IEC 61000-4-8	30 A/m 50Hz or 60Hz	30 A/m 50Hz or 60Hz	Power frequency magnetic field should be at levels characteristic of a typical location in a typical commercial or hospital environment.
Note: UT: rated voltage(s); E.g. 25/30 cycles means 25 cycles at 50Hz or 30 cycles at 60Hz			

Guidance and manufacturer's declaration – electromagnetic immunity			
The CuringPen is intended for use in the electromagnetic environment specified below. The customer or the user of the CuringPen should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted disturbances induced by RF fields IEC 61000-4-6	3 V 0.15 MHz – 80 MHz, 6 V in ISM bands between 0.15 MHz and 80 MHz, 80 % AM at 1 kHz	3 V	Portable and mobile RF communications equipment should be used no closer to any part of the CuringPen , including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.
Radiated RF EM fields IEC 61000-4-3	3 V/m, 80 MHz – 2,7 GHz, 80 % AM at 1 kHz	3V/m	
Proximity fields from RF		Complies	

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd!
 Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

wireless communication equipment IEC 61000-4-3	See the RF wireless communication equipment table in "Recommended minimum separation distances"		Recommended minimum separation distances See the RF wireless communication equipment table in "Recommended minimum separation distances"
--	---	--	--

Recommended minimum separation distances

Nowadays, many RF wireless equipments have being used in various healthcare locations where medical equipment and/or systems are used. When they are used in close proximity to medical equipment and/or systems, the medical equipment and/or systems' basic safety and essential performance may be affected. The **CuringPen** has been tested with the immunity test level in the below table and meet the related requirements of IEC 60601-1-2:2014. The customer and/or user should help keep a minimum distance between RF wireless communications equipments and the **CuringPen** as recommended below.

Test frequenc y (MHz)	Band (MHz)	Service	Modulati on	Maximu m power (W)	Distance (m)	Immunity test level (V/m)
385	380-390	TETRA 400	Pulse modulati on 18Hz	1.8	0.3	27
450	430-470	GMRS 460 FRS 460	FM ± 5 kHz deviation 1 kHz sine	2	0.3	28
710	704-787	LTE Band 13, 17	Pulse modulati on 217Hz	0.2	0.3	9
745						
780						
810	800-960	GSM 800/900,		2	0.3	28
870						

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.

930		TETRA 800, iDEN 820, CDMA 850, LTE Band 5	Pulse modulati on 18Hz			
1720	1700- 1990	GSM 1800; CDMA 1900; GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS	Pulse modulati on 217Hz	2	0.3	28
1845						
1970						
2450	2400- 2570	Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7	Pulse modulati on 217Hz	2	0.3	28
5240	5100- 5800	WLAN 802.11 a/n	Pulse modulati on 217Hz	0.2	0.3	9
5500						
5785						

13.Statement

Service Life

The service life of CuringPen series products is 3 years.

Disposal

The package should be recycled. Metal parts of the device are disposed as scrap metal. Synthetic materials, electrical components, and printed circuit boards are disposed as electrical scrap. The lithium batteries are disposed as special refuse. Please deal with them according to the local environmental protection laws and regulation.

Rights

All rights of modifying the product are reserved to the manufacturer without further notice. The pictures are only for reference. The final interpretation rights belong to CHANGZHOU SIFARY MEDICAL TECHNOLOGY CO., LTD. The industrial design, inner structure, etc, have claimed for several patents by SIFARY, any copy or fake product must take legal responsibilities.

Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić. Błąd! Użyj karty Narzędzia główne, aby zastosować 标题 1 do tekstu, który ma się tutaj pojawić.



Changzhou Sifary Medcial Technology Co., Ltd

Add: NO.99, Qingyang Road, Xuejia County, Xinbei District,
Changzhou City, 213000 Jiangsu, China

Tel: +86-0519-85962691

Fax: +86-0519-85962691

Email: ivy@sifary.com

Web: www.sifary.com

Version: S01

IFU-6535001

Issued: 2019.9.17

All rights reserved.